



# Certificate of Calibration

Certificate Number: 203247

**CUSTOMER:**

**Power Container Corp.**

33 Schoolhouse Rd. Somerset, NJ 08873

Phone (732) 560-3655 | Cell (551)689-6631

Department: Quality

**Calibration Technician:** Sherif Taher

**Remarks:**

**Calibration Procedure Used:** ASME B89. 1.14

USQC Measurement Dept. certifies that the above listed instrument has been calibrated to the manufacture's specifications if available, and/or to. limits specified by customer, using standards traceable to the National Institute of Standards and Technology or which have been derived from accepted values of natural physical constants or through the use of the ratio method of self-calibration techniques.

Methods used are in accordance with FDA(GMP), MIL-STD 45662/A, ANSI Z-540-1994, ISO 17025 and ISO 10012-1. Per ANSI Z-540-1 total uncertainty ratios exceed the minimum 4:1 requirement unless otherwise noted in the report of calibration.

Approved By: Sherif Taher

Date: 4/12/2021



Measurement Dept.

Report of Calibration

Date Arrived: 4/12/2021

Cal Date: 4/12/2021

Cal Due: 4/11/2022

Environmental Conditions: 70.0°F, 40 % RH

#ID	Manufacturer	Serial #	Model	Height Gauge	Digital Force gauge	Test Result
03 17 10 01	Chatillon	03 17 10 01	LTCM-500	Pass	Pass	Pass

Remarks:

- 1- Device in good condition
- 2- Tolerances in acceptable range

Approved By: Sherif Taher

Date: 4/12/2021